

H3800 Hematology Analyzer



Features

Throughput: 60 samples/hour

Touch screen: 10.4 inch.

Detection Channel: Dual-channel for counting

Principle: Pulse baseline tracking technology, Waveform discrimination technology, Intelligent floating threshold technology

Storage: 50,000 sample results, including parameter, histogram and patient information.

3 histograms, alarm for abnormal erythrocytes, leukocyte and platelet.

Sample Volume: Whole blood 9 μ L, Prediluted 20 μ L

Counting modes: Venous whole blood, capillary whole blood and pre-diluted.

Input/output: 4 USB ports and Bi-directional LIS connection.

Printout: Available for thermal printer and USB external printer, various types of printer. Multiple print templates available

Technical specification

Parameters

The detection items include 21 parameters: report parameters (WBC, Lym#, Mid#, Gran#, Lym%, Mid%, Gran%, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-CV, RDW-SD, PLT, MPV, PDW, PCT, P-LCR, and P-LCC), 3 histograms, alarm for abnormal erythrocytes, leukocyte and platelet.

Performance

Parameters	Repeatability	Linearity range:	Carryover
WBC	≤2.0%	(0.00-300.00)X10 ⁹ /L	≤ 0.5%
RBC	≤1.5%	(0.00-8.50)X10 ¹² /L	≤ 0.5%
HGB	1.5%	(0-250) g/L,	≤ 0.5%
PLT	≤4.0%	(0-3000)X10 ⁹ /L	≤ 0.5%
MCV	≤1.0%	0.0-67.0%	≤ 1.0%

Parameter (testing range)	Whole blood precision (CV)	Pre-diluted precision (CV)
WBC (7 - 15.0 x 10 ⁹ /L)	≤ 2.0%	≤ 4.0%
RBC (3.5 - 6.0 x 10 ¹² /L)	≤ 1.5%	≤ 3.0%
HGB (110 - 180 g/L)	≤ 1.5%	≤ 3.0%
MCV (70 - 120 fL)	≤ 1.0%	≤ 2.0%
PLT (150 - 500 x 10 ⁹ /L)	≤ 4.0%	≤ 8.0%

Operation environment: Temperature 10℃to 30℃; relative humidity ≤ 85%

Power supply: 100V-240V; 50/60Hz

Dimension: 364mm(L) x417mm(W) x 477mm(H)

Net weight: ≤ 25KG



Shenzhen Heto Medical Tech Co., Ltd.

Address: 4/F, Building No.4, Shenzhen Environmental Protection Industrial Park,
No.31 Makan South Road, Nanshan District, Shenzhen, 518055, P.R.China.

Tel: +86 755-28935967

Fax: +86 755-28937587

Email: Sales@hetomedical.com

Web: www.hetomedical.com



Contact Us

Local Distributor